Electron Bolus Creation

An electron bolus is an optional device for each electron treatment beam. A user is able to add an electron bolus to their beams to help improve the dose conformality to the distal target surface. The four options for bolus types in the decimal eRT app are fully defined below.

▼ Bolus			
Bolus Type:	None	•	

Fig. 1: Beams Block

Optimized Thickness Bolus

The first option a user has for creating a bolus for the selected beam is an optimized thickness bolus. The goal for an optimized bolus is to best fit the specified Planning Isodose level of the computed beam dose to the distal surface of the target structure. The application provides reasonable default parameters for designing such a bolus, so that in most cases, the user can simply click the "Generate Bolus" button and the application will design a suitable bolus automatically. Once the application has finished calculating, the bolus will get added to the beam.



Fig. 2: Beams Block

For complex and challenging plans, a user can also access advanced options to modify an existing bolus or create a new one from scratch.

In the advanced options you can view and manage various operators to augment the design of the bolus. The following operators are available:

Construction Operators (only available when first creating a new bolus):

• Automated Marching

 This operator constructs the proximal bolus surface using a series of incremental steps. Each step consists of a small thickness reduction at the optimally selected locations, followed by a hot spot damping step, in which the slopes of the bolus surface are reduced in areas where high dose values are found.

- Geometric Sequence
 - This selection performs the following sequence of operators: Create, Smooth, Truncate.

• Single shift Sequence

 $\circ\,$ This selection performs the following sequence of operators: Create, Smooth, Isodose Shift, Smooth, Truncate.

• Double Shift Sequence

 This selection performs the following sequence of operators: Create, Smooth, Isodose Shift, Smooth, Isodose Shift, Smooth, Truncate.

• Create

 This operator constructs the proximal bolus surface by computing the difference between the physical depth to the distal PTV surface and the depth of the Planning Isodose level along each ray line of the bolus construction grid (for ray lines not intersecting the PTV, an extension is performed using the heights of the points that do intersect, which helps ensure the PTV edges are adequately covered).

Modification Operators (only available after a construction operator has been performed):

- Smooth
 - This operator performs a Gaussian type smoothing on the current bolus proximal surface, which uses a distance weighted average of the neighboring point heights to compute a new height for each bolus surface point.
- Isodose shift
 - This operator updates the bolus proximal surface by using the dose results with the current bolus in place to adjust the bolus heights by the difference between the position of the Planning Isodose contour and the position of the distal PTV surface.
- Truncate
 - This operator updates a bolus proximal surface by reducing the height of the bolus in the region outside the Block Outer Border. The height is reduced to be slightly above the highest point in the modulated region. It should be mentioned that truncating a bolus may limit the ability to shift points upward (increase in thickness) when performing manual isodose shift operators after the truncate.

• Specified Shift

• This operator updates a bolus proximal surface by shifting all points (even the region beyond the Block Outer Border) by the specified distance.

• Intensity Modulation

 This operator manually adds an Intensity Modulator device to the beam. If multiple Intensity Modulation operators are added in the operator list, then the results of the multiple iterations of the modulation calculation are combined to produce a single Intensity Modulator device for the beam.

If an operator is selected from the list the UI will update to show the options for that operation. Once the options have been finalized the user can add the operator to the list which will re-compute the bolus and IM devices for the beam.

Below is an example of adding the automated marching to a new bolus:



Fig. 3: New bolus operator options

▼ Bolus				
Bolus Type: Optim	ized Thickness 🔹			
Bolus Material: Blue V	Vax 💌			
External Structure:	Use course external v skin (External)			
Add New Operator:	Automated Marching			
Bolus Geometry Options				
Block Outer Border:	1 cm			
Target Inner Border:	0.2 cm			
Depth Beyond Target:	0.5 cm			
Minimum Thickness:	0.2 cm			
Distal PTV Smoothing:	0 cm			
Include Intensity Mo	dulator			
 Bolus Design Hist 	ory			
	Generate Bolus			
	Hide Advanced Bolus Options			

Fig. 4: Adding an automated marching operator

The advanced optimized thickness bolus design parameters allow the user to fine tune the bolus design for specific operators. These options include:

• Block Outer Border:

 $^{\circ}\,$ The distance to extend the bolus surface beyond the block/collimator perpendicular to the CAX.

• Target Inner Border:

 Specifies the distance to offset the target projection inward when defining the region over which initial construction and smoothing operators act. This is used to reduce the effects that the rapid PTV height changes near the target edges can have on the construction and smoothing operators.

• Depth Beyond Target:

• The beams-eye-view distance past the selected target (along the CAX direction) at which the patient-side surface will be truncated.

• Minimum Thickness:

- The minimum allowable thickness of the bolus as measured along a beam aligned rayline.
- Distal PTV Smoothing:
 - The smoothing ball radius used to smooth the target structure. A larger radius will result in more smoothing but will also cause a greater deviation of the target from the original structure. Note that setting this value to 0 mm will result in no smoothing for the target structure.
- Smoothing Weight: (Smoothing operator only)
 - The exponential smoothing weight coefficient (smaller value increases smoothing)
- Smoothing Size: (Smoothing operator only)
 - The smoothing distance factor (larger values spread the smoothing range further)
- Additional Thickness: (Specified Shift operator only)
 - The thickness to add to the bolus proximal surface (use a negative value to reduce height)

Additionally by selecting the "Bolus Design History" you can see a list of every step in the bolus design. The last step is removable individually, and there is also the option to clear all steps and start from scratch.

▼ Bolu	s Design History		
	0: create		
	1: smooth		
	2: isodose_shift	×	
		clear all	

Fig. 5: Bolus design history list

Uniform Thickness Bolus

decimal App Documentation - http://apps.dotdecimal.com/

Uniform Thickness Bolus devices are designed at the plan level, but they must be added to each intended beam individually to ensure the computed dose accounts for the presence of the bolus in the beam path.

▼ Bolus		
Bolus Type:	Uniform Thickness	•
Plan Level Bolu	us: ensity Modulator	•

Fig. 8: Beams Block

Structure as Bolus

The user also has the option to use existing plan structure as a bolus. Once this option is selected, the user will be able to choose the structure from the drop down list and the app will set it as a bolus for this beam.

Doldo Typo.		J
Structure:	Electron Bolus 🔹	
Include Inte	РТV	
	PTV 6840	
	Rt Parotid	
	Cord	
	Cord+5	
	Ant avoid	
	Post avoid	
	Rt Eye	
	Lt Eye	
	External	
	BOLUS 1CM	
	ptv p5	
	ptv p20	
	nose	
	90 % (Trial_1)_1	
	Electron Bolus	
	~	

Fig. 9: selecting from the structure list

Bolus Type: Use Structure As Bolus 🔻	
Structure: Electron Bolus 🔻	
Include Intensity Modulator	

Fig. 10: selecting a Bolus as a structure

Once the structure is selected the bolus will be added to this beam and dose re-calculated.

Note: The bolus is displayed in the BEV for this beam even though the "Electron Bolus" structure is hidden through the right hand side controls. This shows that the displayed bolus is added as a bolus on this beam and not appearing as a structure.



Fig. 11: Bolus in BEV

If you wish to import a bolus that is not currently in your structure list please refer to the DICOM Patient Import section of the user guide for how to re-import a structure set.

Intensity Modulator Device

Additionally if any type of bolus is added to a beam the user is able to also include an intensity Modulator by selecting the option below any bolus type. The application with calculate the device and display it in the beams UI.

NOTE: In order to add an IMET device the selected beam MUST have a valid bolus.



Fig. 12: BEV IMET

get a better IM image once that actually works...

As with a bolus, the IM device can be hidden and un-hidden from the display using the beam controls on the right hand size.



Fig. 14: RHS options to hide/show devices

