

Optimization Constraints

About Constraints

Constraints can be specified at various levels (*Plan*, *Fraction Group*, *Target/Beam Set*) with Astroid and they will affect different groups of beams depending on their level. *Constraints* at the *Plan* level are applied to the total dose resulting from all beams. *Constraints* at the *Fraction Group* level are applied to the total dose resulting from only the beams in the current *Fraction Group*. *Constraints* at the *Target/Beam Set* level are split evenly and applied individually to each *Beam Set*. In other words, the *Constraint* dose is divided by the number of *Beam Sets* in the *Target*, and this dose is then applied as a constraint to each *Beam Set*, so that either SFO and IMPT can be achieved (see [Fraction Groups](#)). The section below will provide a walk through of the different levels and how constraints are applied at each one.

It should be noted that all constraints are considered “hard limits”- values that must be achieved. *Constraints* drive the feasibility calculation- whether the plan is achievable and should be used to ensure certain minimal clinical parameters are met.

The following constraint types are available. Note certain constraints are available only for *Target* type structures.

- **Min:** The minimum dose the structure must receive
- **Max:** The maximum dose the structure may receive
- **Min Mean:** The minimum mean dose a structure must receive
 - This will drive the dose up across the structure
- **Max Mean:** The maximum mean dose a structure may receive
 - This will limit the mean dose across the structure


The user can choose to apply one or multiple of these constraints to any number of structure.

Working with Constraints


Working with Fraction Group Constraints

Constraints at the *Fraction Group* level apply to the total dose across all beams in the *Fraction Group*

1. Select the *Fraction Group* if it has been created or create a new by selecting *Create New Fraction Group*



2. Choose the target, the phase and number of fractions to be treated within this Fraction Group



3. Choose from the drop down the structure or structures to which constraints should be added

4. Define what constraint(s) should be applied to each structure by choosing the constraint and entering the dose

Edit Fraction Group

Color:

Description:

Phase: orig walkthrough ▾

of Fractions: 44 Total: 44, Remaining: 44

Group Constraints:

● PTV_7920 ✕

Min: add statement

Max: 73 Gy(RBE) ✕

Min Mean: add statement

Max Mean: add statement

Add Structure ▾

Target List

Add Target

5. Once the constraints are set in the *Fraction Group* the user will create *Target Lists* and assign *Constraints*
6. The assigned constraint doses will be divided evenly among the *Beam Sets* in the *Target List*

Target List

Add Target

Target: PTV_7920 ▾

Beam Sets:

▾ beam set remove

b1 - G90 C0; 12 cm; no shifter; ✕

add beam ▾

▾ beam set remove

b2 - G270 C0; 12 cm; no shifter; ✕

add beam ▾

add beam set

Constraints:

● PTV_7920 ✕

Min: 70 Gy(RBE) ✕

Max: add statement

Min Mean: add statement

Max Mean: add statement


Add Structure ▾

Done Cancel


Working with Plan Constraints

Constraints at the *Plan* level are applied to the total dose across all beams.


1. Open the *Constraint* sub block contained in the *Plan Constraints/Objectives* block and choose the *Edit* button.



2. Choose from the drop down the structure or structures to which constraints should be added



3. Define what constraint(s) should be applied to each structure by choosing the constraint and entering the dose



4. Follow this and enter the constraints for all applicable structures.

Plan Constraints / Objectives

Constraints

- PTV_7920** (Red dot)
- Min: c_y(PPTZ)
- Max: c_y(PPTZ)
- Min Mean:
- Max Mean:
- Rectum** (Yellow dot)
- Max:
- Max Mean: c_y(PBEZ)
- Bladder** (Yellow dot)
- Max:
- Max Mean: c_y(PBEZ)

Add Structure

5. When finished click the OK button.

6. Once all the Constraints have been set the user can either start the Feasibility or move on to defining the Objectives

From:
<http://apps.dotdecimal.com/> - **decimal App Documentation**

Permanent link:
http://apps.dotdecimal.com/doku.php?id=planning:userguide:tutorials:optimization_constraints&rev=1487194627

Last update: **2021/07/29 18:24**

