

# Overview

The .decimal astroid Dosimetry App device is used for planning and analysis of proton radiation therapy treatments. The astroid Dosimetry App device is not an interactive end user application. Users of the system will write scripts or use a fully interactive software program that makes calls to the functions provided by the astroid Dosimetry App. The core functionality includes various CT image processing tools, structure and contour modification operators, proton dose calculations, proton aperture and range compensator device design algorithms, and many other radiotherapy specific calculation functions.

Access to the astroid Dosimetry App is provided by the thinknode™ framework using http json formatted requests. thinknode™ provides the 'backbone' used to send and receive requests, maintain users, realms and organizations, and provide data storage and management.

## User Guide

The Dosimetry App [user guide](#) lists all available api function calls, as well as gives examples of usage and explanation of the affects.

**Getting Started** Connecting to thinknode™ api and initial setup of the astroid Dosimetry App.

**Proton Delivery System** Details about coordinate systems and beam representations.

**Function Categories** Dose Calculation Functions (DCF), Design Task Functions (DTF), and Radiotherapy Support Functions (RSF) available through the thinknode™ api.

**Examples** thinknode™, DCF, DTF, and RSF example projects and usages.

**Known Limitations** Known application limitations, defects, or inconsistencies.

**Data Types** astroid Dosimetry App datatypes available through the thinknode™ api.

## Commissioning Guide

The Dosimetry App [Commissioning Guide](#) outlines and walks users through the process of creating a machine model of a clinical proton therapy beamline.

**Getting Started** Setup and data collection.

**Testing Responsibilities** Here's what you have to test before clinical use can begin

**Data Requirements** Lists and describes the data needed for commissioning a proton beam model.

**Examples** Here's an example machine model

# Instructions For Use

The Dosimetry App [Instructions For Use](#) outlines the intended use and user requirements of using the Dosimetry App.

**Overview** Intended use and indications for use of **User Profile** Recommended user education and the application. experience level.

**Warning** Warning of potential misuse.

**Testing Responsibilities** Testing responsibilities for ensuring correct setup and configuration of the astroid Dosimetry App.

**Product Features** High level features of the astroid Dosimetry App.

## Reference Documentation

**Hong et al** A pencil beam algorithm for proton dose calculations

**Slopsema** Incorporation of the aperture thickness in proton pencil-beam dose calculations

**IAEA-TRS-430** Commissioning and Quality Assurance of Computerized Planning Systems for Radiation Treatment of Cancer

**IAEA-TECDOC-1583** Commissioning of Radiotherapy Treatment Planning Systems: Testing for Typical External Beam Treatment Techniques

**Park Commissioning** Commissioning a Proton Therapy Machine and TPS

From:

<http://apps.dotdecimal.com/> - **decimal App Documentation**

Permanent link:

<http://apps.dotdecimal.com/doku.php?id=start&rev=1424273212>

Last update: **2021/07/29 18:18**

